

[Cat. No.] **K-2908**

Introduction

AccuPower® Leptospira PCR Kit is a ready-to-use premix for PCR that can be used to detect the infection of *Leptospira*, a genus of pathogenic spirochete bacteria.

Leptospira is a bacterium that causes acute febrile systemic disease. It is spread by a direct contact with the urine of an infected organism, contaminated water, or an indirect contact with the environment.

Severity of symptoms range from mild to severe, such as headache, muscle pain, fever, and bleeding. *Leptospira* can survive for a long time in a humid environment, which is why acute febrile systemic diseases mainly occur from August to November in South Korea.

This product contains vacuum-dried components specific to *Leptospira* including DNA polymerase, primers, dNTPs, and reaction buffer required for PCR. This ready-to-use kit simplifies preparation of PCR mixture as the user only has to add template DNA and nuclease-free water. After the reaction, since tracking dye is included, the samples can be applied directly on agarose gel for analysis without adding extra solution.

Features & Benefits

- **Convenience & Reproducibility:** All reactants necessary for PCR including primers are lyophilized in each PCR tube, providing reproducible results in a convenient way.
- **Sensitivity:** By applying the patented PyroHotStart (Enzyme-mediated HotStart) technology that minimizes non-specific reactions and maximizes reaction efficiency, only the target gene can be effectively amplified even with a trace amount of template DNA.
- **Stability:** Included stabilizer in the PCR reaction mixture provides increased stability compared to solution-type products.

Composition

Composition	20 µl reaction
Top DNA Polymerase	1 U
dNTPs (dATP, dCTP, dGTP, dTTP)	Each 250 µM
Reaction buffer with 1.5 mM MgCl ₂	1X
Stabilizer and tracking dye	0
Sec-Y Forward primer	0.5 µM
Sec-Y Reverse primer	0.5 µM

* **Note:** For research use only. Not for use in diagnostic or therapeutic procedures.

Specifications

Top DNA Polymerase	
5'→3' exonuclease activity	No
3'→5' exonuclease activity	No
3'-A overhang	Yes
Fragment size	285 bp

Storage

Store at -20°C. If stored in the recommended temperature, this product will be stable until the expiration date printed out on the label.

Online Resources



English

Visit our **product page** for additional information and protocols


Ordering Information

Description	Cat. No.
AccuPower® Leptospira PCR Kit, 0.2 ml thin-wall 8-tube strips with attached cap / 96 tubes	K-2908





Notice

BIONEER corporation reserves the right to make corrections, modifications, improvements and other changes to its products, services, specifications or product descriptions at any time without notice.

Explanation of Symbols



Experimental Procedures

Steps		Procedure Details																								
1	 Add template DNA	<p>1. After preparing the template DNA and nuclease-free water, add the template DNA to the <i>AccuPower[®] Leptospira PCR Kit</i>.</p>																								
2	 Preparation of reaction mixture	<p>2. Add nuclease-free water into PCR tubes to make a total volume of 20 μl. (Do not include the volume of the dried premix in the PCR tubes.)</p> <p>3. Completely dissolve the vacuum-dried pellet by vortexing, and briefly spin down.</p>																								
3	 Incubate reactions in a thermal cycler	<p>4. Place PCR tubes on the thermal cycler.</p> <p>5. Perform the reaction under the following conditions.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: center;">Step</th> <th style="text-align: center;">Temperature</th> <th style="text-align: center;">Time</th> <th style="text-align: center;">Cycles</th> </tr> </thead> <tbody> <tr> <td>Pre-denaturation</td> <td style="text-align: center;">92°C</td> <td style="text-align: center;">5 min</td> <td style="text-align: center;">1 cycle</td> </tr> <tr> <td>Denaturation</td> <td style="text-align: center;">92°C</td> <td style="text-align: center;">30 sec</td> <td></td> </tr> <tr> <td>Annealing</td> <td style="text-align: center;">57°C</td> <td style="text-align: center;">30 sec</td> <td style="text-align: center;">45 cycles</td> </tr> <tr> <td>Extension</td> <td style="text-align: center;">72°C</td> <td style="text-align: center;">30 sec</td> <td></td> </tr> <tr> <td>Final extension</td> <td style="text-align: center;">72°C</td> <td style="text-align: center;">10 min</td> <td style="text-align: center;">1 cycle</td> </tr> </tbody> </table> <p>* Note: Users can adjust the protocol according to their instrument and template sequences to get optimal results.</p>	Step	Temperature	Time	Cycles	Pre-denaturation	92°C	5 min	1 cycle	Denaturation	92°C	30 sec		Annealing	57°C	30 sec	45 cycles	Extension	72°C	30 sec		Final extension	72°C	10 min	1 cycle
Step	Temperature	Time	Cycles																							
Pre-denaturation	92°C	5 min	1 cycle																							
Denaturation	92°C	30 sec																								
Annealing	57°C	30 sec	45 cycles																							
Extension	72°C	30 sec																								
Final extension	72°C	10 min	1 cycle																							
4	 Analyze with gel electrophoresis	<p>6. After the reaction, maintain the reaction mixture at 4-8°C.</p> <p>7. Load samples on agarose gel without adding a loading-dye mixture, and perform gel electrophoresis for analysis.</p>																								