

[Cat. No.] **K-6827**

Introduction

AccuPower[®] Bacteroides thetaiotaomicron Real-Time PCR Kit is a product that can specifically detect *Bacteroides thetaiotaomicron* (*B. thetaiotaomicron*) by real-time PCR.

B. thetaiotaomicron is a gram-negative bacterium that causes peritonitis, sepsis, or suppurative arthritis. *B. thetaiotaomicron* is a part of the intestinal flora residing in the gastrointestinal tract and is an opportunistic pathogen that exhibits pathogenic properties through infection with intestinal wounds. In addition, it is a pathogen that needs to be careful not only because of direct infection to other tissues but also because the substance produced during the metabolic process contributes to the growth of other pathogens. This product contains all Real-time PCR components specific to *B. thetaiotaomicron*, including DNA polymerase, dNTPs, and a reaction buffer. The users can easily prepare a reaction mixture simply by adding template DNA, Oligo Mix, and DEPC-D.W.

Features & Benefits

- Convenience: All necessary reactants for real-time PCR are included in a tube (i.e., Master Mix Type), allowing the users to perform reaction simply by adding template DNA, Oligo Mix, and DEPC-D.W.
- Sensitivity: By using BIONEER's HotStart *Taq* DNA Polymerase that minimizes non-specific reactions and maximizes reaction efficiency, only the target gene can be effectively amplified even with a trace amount of template DNA.

Components

Components	Amount
2X Master Mix	625 µl x 2 ea
Oligo Mix	500 µl
DEPC-D.W.	1.8 ml
Positive Control (1x10 ⁸ copies/µl)	50 µl

* **Note:** For research use only. Not for use in diagnostic or therapeutic procedures.

Composition

Composition	25 µl reaction
2X Master Mix	
<i>Taq</i> DNA Polymerase	2.5 U
dNTPs (dATP, dCTP, dGTP, dTTP)	Each 300 µM
Reaction buffer with 2 mM MgCl ₂	1X
Oligo Mix	
<i>B. thetaiotaomicron</i> Forward primer	0.4 µM
<i>B. thetaiotaomicron</i> Reverse primer	0.4 µM
<i>B. thetaiotaomicron</i> Probe (FAM)	0.4 µM
ROX dye	1X

Specifications

<i>Taq</i> DNA Polymerase	
5'→3' exonuclease activity	Yes
3'→5' exonuclease activity	No
3'-A overhang	Yes

Storage

Store at -20°C. If stored in the recommended temperature, this product will be stable until the expiration date printed out on the label.

Online Resources



English

Visit our **product page** for additional information and protocols


Ordering Information

Description	Cat. No.
<i>AccuPower</i> [®] Bacteroides thetaiotaomicron Real-Time PCR Kit, 1.25 ml of 2X Master Mix solution, 100 tests	K-6827




Notice

BIONEER corporation reserves the right to make corrections, modifications, improvements and other changes to its products, services, specifications or product descriptions at any time without notice.

Explanation of Symbols



Experimental Procedures

Steps		Procedure Details															
1	 Preparation of reaction mixture	1. Thaw all components of <i>AccuPower</i> [®] Bacteroides thetaiotaomicron Real-Time PCR Kit on ice and mix thoroughly before use. Then, briefly spin down all components.															
2	 Composition of reaction mixture	2. Add all components into PCR tubes (not provided) or a plate (not provided) referring to the following list of components (Based on 1 test). <table border="1" style="width: 100%; margin-top: 10px;"> <thead> <tr> <th style="text-align: left;">Components</th> <th style="text-align: left;">Amount</th> </tr> </thead> <tbody> <tr> <td>2X Master Mix</td> <td>12.5 µl</td> </tr> <tr> <td>Oligo Mix</td> <td>5 µl</td> </tr> <tr> <td>Template DNA</td> <td>1-5 µl</td> </tr> <tr> <td>DEPC-D.W.</td> <td>Variable</td> </tr> <tr> <td>Total volume</td> <td>25 µl</td> </tr> </tbody> </table>	Components	Amount	2X Master Mix	12.5 µl	Oligo Mix	5 µl	Template DNA	1-5 µl	DEPC-D.W.	Variable	Total volume	25 µl			
Components	Amount																
2X Master Mix	12.5 µl																
Oligo Mix	5 µl																
Template DNA	1-5 µl																
DEPC-D.W.	Variable																
Total volume	25 µl																
3	 Real-time PCR	3. Place PCR tubes or plate on the Real-Time Quantitative thermal cycler. 4. Perform the reaction under the following conditions. <table border="1" style="width: 100%; margin-top: 10px;"> <thead> <tr> <th style="text-align: left;">Step</th> <th style="text-align: left;">Temperature</th> <th style="text-align: left;">Time</th> <th style="text-align: left;">Cycles</th> </tr> </thead> <tbody> <tr> <td>Pre-denaturation</td> <td>95°C</td> <td>5 min</td> <td>1 cycle</td> </tr> <tr> <td>Denaturation</td> <td>95°C</td> <td>10 sec</td> <td rowspan="2">45 cycles</td> </tr> <tr> <td>Annealing & Extension</td> <td>55°C</td> <td>20 sec</td> </tr> </tbody> </table> <p>* Note: Users can adjust the protocol according to their instrument and template sequences to get optimal results.</p> 5. After the reaction is completed, analyze the results.	Step	Temperature	Time	Cycles	Pre-denaturation	95°C	5 min	1 cycle	Denaturation	95°C	10 sec	45 cycles	Annealing & Extension	55°C	20 sec
Step	Temperature	Time	Cycles														
Pre-denaturation	95°C	5 min	1 cycle														
Denaturation	95°C	10 sec	45 cycles														
Annealing & Extension	55°C	20 sec															