

[Cat. No.] **K-6842**

Introduction

AccuPower® Haemophilus parainfluenzae Real-Time PCR Kit is a product that can specifically detect *Haemophilus parainfluenzae* (*H. parainfluenzae*) by real-time PCR.

H. parainfluenzae is a gram-negative bacterium commonly found in the mouth and pharynx. And it is one of the HACEK organisms that causes endocarditis. Infectious endocarditis may be accompanied by embolism, which may affect the patient's prognosis.

H. parainfluenzae has a lipid- oligosaccharide component with modified surface residues, so it is easy to attach to and colonize tissues. These characteristics allow long-term survival in the host's mucosal cells to sustain infection. For the treatment of *H. parainfluenzae* infection, administering of β -lactam antibiotics is usually common.

This product contains all Real-time PCR components specific to *H. parainfluenzae*, including DNA polymerase, dNTPs, and reaction buffer. The users can easily prepare reaction mixture simply by adding template DNA, Oligo Mix and DEPC-D.W.

Features & Benefits

- Convenience: All necessary reactants for real-time PCR are included in a tube (i.e., Master Mix Type), allowing the users to perform reaction simply by adding template DNA, Oligo Mix, and DEPC-D.W.
- Sensitivity: By using BIONEER's HotStart Taq DNA Polymerase that minimizes non-specific reactions and maximizes reaction efficiency, only the target gene can be effectively amplified even with a trace amount of template DNA.

Components

Components	Amount
2X Master Mix	625 μ l x 2 ea
Oligo Mix	500 μ l
DEPC-D.W.	1.8 ml
Positive Control (1x10 ⁸ copies/ μ l)	50 μ l

* **Note:** For research use only. Not for use in diagnostic or therapeutic procedures.

Composition

Composition	25 μ l reaction
2X Master Mix	Taq DNA Polymerase 2.5 U
	dNTPs (dATP, dCTP, dGTP, dTTP) Each 300 μ M
	Reaction buffer with 2 mM MgCl ₂ 1X
Oligo Mix	<i>H. parainfluenzae</i> Forward primer 0.4 μ M
	<i>H. parainfluenzae</i> Reverse primer 0.4 μ M
	<i>H. parainfluenzae</i> Probe (FAM) 0.4 μ M
	ROX dye 1X

Specifications

Taq DNA Polymerase	
5'→3' exonuclease activity	Yes
3'→5' exonuclease activity	No
3'-A overhang	Yes

Storage

Store at -20°C. If stored in the recommended temperature, this product will be stable until the expiration date printed out on the label.

Online Resources



English

Visit our **product page** for additional information and protocols

Ordering Information

Description	Cat. No.
AccuPower® Haemophilus parainfluenzae Real-Time PCR Kit, 1.25 ml of 2X Master Mix solution, 100 tests	K-6842

Notice

BIONEER corporation reserves the right to make corrections, modifications, improvements and other changes to its products, services, specifications or product descriptions at any time without notice.

Explanation of Symbols



Batch Code



Biological Risks



Catalog Number



Caution



Consult Instructions For Use



Contains Sufficient for <n> tests



Do not Re-use



Manufacturer



Research Use Only






Temperature Limitation



Use-by Date

Experimental Procedures

Steps		Procedure Details																
1	<div></div> <div>Preparation of reaction mixture</div>	1. Thaw all components of <i>AccuPower</i> ® <i>Haemophilus parainfluenzae</i> Real-Time PCR Kit on ice and mix thoroughly before use. Then, briefly spin down all components.																
2	<div></div> <div>Composition of reaction mixture</div>	<div>2. Add all components into PCR tubes (not provided) or a plate (not provided) referring to the following list of components (Based on 1 test).</div> <table><tr><th>Components</th><th>Amount</th></tr><tr><td>2X Master Mix</td><td>12.5 µl</td></tr><tr><td>Oligo Mix</td><td>5 µl</td></tr><tr><td>Template DNA</td><td>1-5 µl</td></tr><tr><td>DEPC-D.W.</td><td>Variable</td></tr><tr><td>Total volume</td><td>25 µl</td></tr></table>		Components	Amount	2X Master Mix	12.5 µl	Oligo Mix	5 µl	Template DNA	1-5 µl	DEPC-D.W.	Variable	Total volume	25 µl			
Components	Amount																	
2X Master Mix	12.5 µl																	
Oligo Mix	5 µl																	
Template DNA	1-5 µl																	
DEPC-D.W.	Variable																	
Total volume	25 µl																	
3	<div></div> <div>Real-time PCR</div>	<div>3. Place PCR tubes or plate on the Real-Time Quantitative thermal cycler.</div> <div>4. Perform the reaction under the following conditions.</div> <table><tr><th>Step</th><th>Temperature</th><th>Time</th><th>Cycles</th></tr><tr><td>Pre-denaturation</td><td>95°C</td><td>5 min</td><td>1 cycle</td></tr><tr><td>Denaturation</td><td>95°C</td><td>10 sec</td><td rowspan="2">45 cycles</td></tr><tr><td>Annealing & Extension</td><td>55°C</td><td>20 sec</td></tr></table> <div>* Note: Users can adjust the protocol according to their instrument and template sequences to get optimal results.</div> <div>5. After the reaction is completed, analyze the results.</div>		Step	Temperature	Time	Cycles	Pre-denaturation	95°C	5 min	1 cycle	Denaturation	95°C	10 sec	45 cycles	Annealing & Extension	55°C	20 sec
Step	Temperature	Time	Cycles															
Pre-denaturation	95°C	5 min	1 cycle															
Denaturation	95°C	10 sec	45 cycles															
Annealing & Extension	55°C	20 sec																