

[Cat. No.] **K-6603**

Introduction

AccuPower® Plus DualStar™ qPCR Master Mix is a ready-to-use mixture for real-time PCR with enhanced specificity and sensitivity by applying hydrolysis probe method and antibody-based HotStart Taq DNA Polymerase. By applying antibody-based HotStart Taq DNA Polymerase, it provides reduced non-specific reactions such as mis-priming and primer dimer during PCR at a low temperature. This product contains all components for real-time PCR, except for template DNA, target-specific primers, and fluorogenic probe. By just adding template DNA, target-specific primers, and probe, reproducible results with high sensitivity and specificity can be obtained. This product can be used for hydrolysis probe-based real-time PCR experiments for the amplification and detection of genomic DNA and cDNA targets, differential gene expression profiling, single nucleotide polymorphism (SNP) analysis, and evaluation of RNAi products.

Applications

- Gene expression profiling
- Target DNA quantification
- Microbial detection
- Viral/bacterial pathogen load determination
- Evaluation of primer pair performance for probe-based real-time PCR

Features & Benefits

- Dynamic range: A wide range of 8 logs up to 10-10⁸ copies.
- Specificity: Optimized amplification of target gene using HotStart Taq DNA Polymerase.
- Comprehensiveness: Effective real-time PCR regardless of gene types, including DNA, cDNA and high GC templates.
- Convenience: Reactants are included in a tube, it allows any user simply perform real-time PCR by adding template DNA, target-specific primers, and probe.
- Reproducibility: Mass production under ISO 9001 quality system allows minimized deviation between lots and reproducible results in replicated tests performed under same conditions and variation.

Components

Components	K-6603
2X Master Mix	0.625 ml x 4 ea
50X ROX dye	0.1 ml
DEPC-D.W.	1.2 ml

* **Note:** ROX dye is used for normalization of intensity by background subtraction. The use of ROX dye is recommended for Applied Biosystems 7500 Real-Time PCR System (Applied Biosystems), but not required for Exicycler™ 96 Real-Time PCR System (BIONEER) and CFX96 Real-Time PCR System (Bio-Rad).

Composition

2X Master Mix	Concentration
HotStart Taq DNA Polymerase	1 U
dNTPs (dATP, dCTP, dGTP, dTTP)	1.2 mM
Reaction buffer with 2 mM MgCl ₂	1X
Stabilizer	1X

Specifications

HotStart Taq DNA Polymerase	
5' to 3' exonuclease activity	Yes
3' to 5' exonuclease activity	No
3'-A overhang	Yes

Storage

Store at -20°C. If stored in the recommended temperature, this product will be stable until the expiration date printed out on the label.

Online Resources



Korean



English

Visit our **product page** for additional information and protocols

Ordering Information

Description	Cat. No.
2.5 ml of 2X Master Mix solution	0.625 ml x 4 ea K-6603

Notice

BIONEER corporation reserves the right to make corrections, modifications, improvements and other changes to its products, services, specifications or product descriptions at any time without notice.

Explanation of Symbols



Batch Code



Biological Risks



Catalog Number



Caution



Consult Instructions For Use



Contains Sufficient for <n> tests



Do not Re-use



Manufacturer



Research Use Only






Temperature Limitation



Use-by Date

Experimental Procedures

Steps		Procedure Details																		
1	 Thaw reagents	<p>1. Thaw <i>AccuPower® Plus DualStar™</i> qPCR Master Mix on ice and mix thoroughly before use. Then, briefly spin down components.</p> <p>2. Dispense appropriate volumes of <i>AccuPower® Plus DualStar™</i> qPCR Master Mix into PCR tubes (not provided). Use 25 µl of 2X Master Mix for 50 µl reaction.</p>																		
2	 Preparation of reaction mixture	<p>3. Add template DNA, target-specific primers, hydrolysis probe (not provided), 50X ROX dye (optional), and DEPC-D.W. into PCR tubes containing <i>AccuPower® Plus DualStar™</i> qPCR Master Mix.</p> <ul style="list-style-type: none"> Preparation of reaction mixture <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: center;">Components</th> <th style="text-align: center;">50 µl reaction</th> </tr> </thead> <tbody> <tr> <td>2X Master Mix</td> <td style="text-align: center;">25 µl</td> </tr> <tr> <td>Template DNA (10 pg-100 ng)</td> <td style="text-align: center;">Variable</td> </tr> <tr> <td>Forward primer (10 pmol/µl)</td> <td style="text-align: center;">0.5-5 µl</td> </tr> <tr> <td>Reverse primer (10 pmol/µl)</td> <td style="text-align: center;">0.5-5 µl</td> </tr> <tr> <td>Hydrolysis probe (10 pmol/µl)</td> <td style="text-align: center;">0.5-5 µl</td> </tr> <tr> <td>(Optional) 50X ROX dye</td> <td style="text-align: center;">1 µl</td> </tr> <tr> <td>DEPC-D.W.</td> <td style="text-align: center;">Variable</td> </tr> <tr> <td>Total volume</td> <td style="text-align: center;">50 µl</td> </tr> </tbody> </table> <p>* Note: This protocol was validated with the TaqMan® probe as a hydrolysis probe.</p> <p>4. Seal real-time PCR tubes or plate with adhesive optical sealing film (Cat. No. 3111-4110, not provided).</p> <p>5. Mix the reaction mixture by vortexing, and briefly spin down.</p>	Components	50 µl reaction	2X Master Mix	25 µl	Template DNA (10 pg-100 ng)	Variable	Forward primer (10 pmol/µl)	0.5-5 µl	Reverse primer (10 pmol/µl)	0.5-5 µl	Hydrolysis probe (10 pmol/µl)	0.5-5 µl	(Optional) 50X ROX dye	1 µl	DEPC-D.W.	Variable	Total volume	50 µl
Components	50 µl reaction																			
2X Master Mix	25 µl																			
Template DNA (10 pg-100 ng)	Variable																			
Forward primer (10 pmol/µl)	0.5-5 µl																			
Reverse primer (10 pmol/µl)	0.5-5 µl																			
Hydrolysis probe (10 pmol/µl)	0.5-5 µl																			
(Optional) 50X ROX dye	1 µl																			
DEPC-D.W.	Variable																			
Total volume	50 µl																			
3	 Real-time PCR	<p>6. Perform the reaction under the following conditions.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: center;">Step</th> <th style="text-align: center;">Temperature</th> <th style="text-align: center;">Time</th> <th style="text-align: center;">Cycles</th> </tr> </thead> <tbody> <tr> <td>Pre-denaturation</td> <td style="text-align: center;">95°C</td> <td style="text-align: center;">3-5 min</td> <td style="text-align: center;">1 cycle</td> </tr> <tr> <td>Denaturation</td> <td style="text-align: center;">95°C</td> <td style="text-align: center;">5-30 sec</td> <td rowspan="2" style="text-align: center;">40-45 cycles</td> </tr> <tr> <td>Annealing & Extension</td> <td style="text-align: center;">55-60°C</td> <td style="text-align: center;">30-35 sec</td> </tr> </tbody> </table> <p>* Note: Users can adjust the protocol according to their instrument and template DNA sequences to get optimal results.</p> <p>7. After the reaction is completed, analyze the results.</p>	Step	Temperature	Time	Cycles	Pre-denaturation	95°C	3-5 min	1 cycle	Denaturation	95°C	5-30 sec	40-45 cycles	Annealing & Extension	55-60°C	30-35 sec			
Step	Temperature	Time	Cycles																	
Pre-denaturation	95°C	3-5 min	1 cycle																	
Denaturation	95°C	5-30 sec	40-45 cycles																	
Annealing & Extension	55-60°C	30-35 sec																		